

Xmaru1717 Image Acquisition Time Measure

- **A Brief**

The test is designed to clarify the acquisition time of the Xmaru1717 Panel for each condition.

- **Requirement for test**

Xmaru1717-1Set, Generator-1Set, Console PC – 1Set, VADAV Library

- **The way of test**

1. Measure and writing the test result which the process time for proceed to calibration and without calibration and averaging all of data for each case.

- **Result Table**

	Without Calibration	Offset Calibration	Offset, Gain Calibration	Calibration
1	5.61	5.42	6.4	6.97
2	5.61	5.83	6.4	6.22
3	5.62	5.41	6.57	6.2
4	5.63	6	6.23	6.19
5	5.42	5.62	6.41	6.4
6	5.42	5.81	6.39	6.39
7	5.61	5.83	6.02	6.19
8	5.62	5.61	6.39	6.19
9	5.62	5.82	6.4	6.4
10	5.83	6.02	6.2	6.2
Average	5.59	5.73	6.34	6.33

